

# Honeywell

CERTIFICATE OF ANALYSIS  
Page 1 of 1

Honeywell Specialty Chemicals Seelze GmbH D-30918 Seelze  
Telefon: +49 5137 999-353

501605  
ALD-34840-50

10.06.2024 / COAH\_01111918

Analyzed for:  
Honeywell PMT Advanced Materials E&C  
Offering Management  
115 Tabor Road, Morris Plains  
NJ 07950, USA

QC Release Date: 10.Jun.2024  
Production date: 03.May.2024  
Expiry Date: 07.Apr.2029

Article/Product: 34840  
HYDRANAL-Coulomat CG

Lot: O124G

The quality certificate is valid for the time of delivery

HYDRANAL-Coulomat CG is a catholyte for coulometric Karl Fischer titration (methanol based).

Parameter	Specification	QC results
Suitability (for KF-Titration)*	test	test passed
Assay of water*	max. 50 ppm	10 ppm
Number of samples (n)	minimum 3	5

\*coulometric KF titration according to Test Method 02 HYDRANAL version 13.Nov.2017  
The KF suitability test was performed on an original ampoule from the filling line.  
The assay of water is the highest value from a series of measurements on (n) samples.

- The expiry date is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.
- We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.
- Particular properties of the products or the suitability for a particular area of application are not assured.
- We guarantee a proper quality within our General Conditions of Sales.

*Thomas Wendt*

Honeywell Specialty Chemicals Seelze GmbH  
Thomas Wendt  
Supervisor HYDRANAL Center of Excellence  
Wunstorfer Str. 40, D-30926 Seelze



Quality Testing Lab Accredited According to DIN EN ISO/IEC 17025

