

502017
HA-23864-42



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

Certificate of Analysis

PRODUCT: Amino Acid F Reagent Solution

PRODUCT NUMBER: 2386442

LOT NUMBER: A4347

MANUFACTURE DATE: 12/30/2024

DATE OF ANALYSIS: 12/30/2024

TEST	SPECIFICATIONS	RESULTS
Absorbance of Silica standard vs the reagent blank with 2.0 ppm SiO ₂ .	0.692 to 0.718 abs	0.6960 abs
Reagent Blank: Measured at 815 nm in 1 cm cell vs DI water.	0 to 0.005 abs	0.0011 abs
pH of the solution	7.1 to 7.7	7.57

The expiration date is Aug 2025

Certified by: *Scott Als*

Analytical Services Chemist