



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

503405  
HA-1117-29

### *Certificate of Analysis*

**PRODUCT:** Silica Standard Solution, 50 mg/l as SiO<sub>2</sub>

**PRODUCT NUMBER:** 111729

**LOT NUMBER:** A4305

**MANUFACTURE DATE:** 11/11/2024

**DATE OF ANALYSIS:** 11/11/2024

TEST	SPECIFICATIONS	RESULTS
Silica Concentration	49 to 51 ppm	50.0 ppm

The expiration date is Nov 2025

The item 111729 is traceable to NIST standards SRM 3150 Silicon 10 mg/mL LOT 130912

Certified by: *Scott Als*

Analytical Services Chemist