

500612  
HA-1995-49



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

## Certificate of Analysis

**PRODUCT:** Molybdate 3 Reagent for Silica

**PRODUCT NUMBER:** 199549

**LOT NUMBER:** A4305

**MANUFACTURE DATE:** 11/12/2024

**DATE OF ANALYSIS:** 11/08/2024

TEST	SPECIFICATIONS	RESULTS
Concentration of a treated 1.0 ppm Silica standard measured at 815 nm in a one inch cell.	0.92 to 1.08 ppm	0.978 ppm
Series 5500 Blank (ppb)	0 to 5 ppb	2.2 ppb
31201 Instrument Blank (ppb)	0 to 5 ppb	1.7 ppb
Lab Blank (ppb)	0 to 6 ppb	2.0 ppb

The expiration date is Nov 2029

Certified by: *Scott Als*

Analytical Services Chemist