



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

503402
HA-854-28

Certificate of Analysis

PRODUCT: FerroVer® Iron Reagent

PRODUCT NUMBER: 85428

LOT NUMBER: A4304

MANUFACTURE DATE: 11/11/2024

DATE OF ANALYSIS: 11/11/2024

TEST	SPECIFICATIONS	RESULTS
Full Scale Test: Absorbance of a 5+ ppm iron standard measured at 510 nm vs the reagent blank.	1.45 to 4.2 abs	3.445 abs
Function Test: Absorbance of a 1.00 ppm iron standard measured at 510 nm vs the reagent blank.	0.45 to 0.524 abs	0.4940 abs
Reagent Blank: Measured at 510 nm vs deionized water in 1 inch cells.	0 to 0.009 abs	0.0010 abs

The expiration date is Nov 2028

Certified by: *Scott Als*

Analytical Services Chemist