

503412
HA-21258-15



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

PRODUCT: Digestion Solution for COD 3-150 mg/L Range

PRODUCT NUMBER: 2125815

LOT NUMBER: A4345B

MANUFACTURE DATE: 01/02/2025

DATE OF ANALYSIS: 01/07/2025

TEST	SPECIFICATIONS	RESULTS
Vial to Vial Variability: Statistical sampling of reagent blank absorbance vs DI water, measured at 420 nm and controlled to a 99% confidence level.	0 to 15 ppm	6.0 ppm
Function Test: Determination of the COD in a 120 ppm COD standard prepared from KHP vs a reagent blank.	114 to 126 ppm	124.0 ppm

The expiration date is Dec 2029

Certified by: *Scott Als*

Analytical Services Chemist