

500634
HA-67749-02



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

Certificate of Analysis

PRODUCT: 5500sc Citric Acid-Surfactant Solution, Silica

PRODUCT NUMBER: 6774902

LOT NUMBER: A4298

MANUFACTURE DATE: 11/13/2024

DATE OF ANALYSIS: 11/14/2024

TEST	SPECIFICATIONS	RESULTS
Function, 1 mg/l silica standard	0.325 to 0.375 abs	0.3509 abs
Reagent blank, DI water.	0 to 0.004 abs	0.0000 abs
Foam is present.	To Pass	Passed
Reagent blank using 10 mg/l phosphate solution.	0 to 0.006 abs	0.0007 abs

The expiration date is Oct 2026

Certified by: *Scott Als*

Analytical Services Chemist