



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500629

HA-26601-53

Certificate of Analysis

PRODUCT: StablCal® Standard, 20 NTU

PRODUCT NUMBER: 2660153

LOT NUMBER: A4290

MANUFACTURE DATE: 11/04/2024

DATE OF ANALYSIS: 11/07/2024

TEST	SPECIFICATIONS	RESULTS
Turbidity	19 to 21 NTU	20.2 NTU

The expiration date is Oct 2026

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by: *Scott Als*

Analytical Services Chemist