



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500905

HA-26087-99 *Certificate of Analysis*

PRODUCT: TPTZ Iron Reagent

PRODUCT NUMBER: 2608799

LOT NUMBER: A4317

MANUFACTURE DATE: 11/20/2024

DATE OF ANALYSIS: 11/20/2024

TEST	SPECIFICATIONS	RESULTS
Absorbance of a 1 mg/L Iron standard vs. the reagent blank.	0.863 to 0.929 abs	0.9210 abs
pH of two pillow fill dissolved in 20 mL of 1 mg/L iron standard.	3.1 to 3.5	3.17
Absorbance of a 1.8 mg/L Iron standard vs. the reagent blank.	1.5 to 1.65 abs	1.526 abs
Reagent blank: Absorbance measured at 590 nm in 1 inch cells vs. DI water.	0 to 0.027 abs	0.0080 abs

The expiration date is Mar 2028

Certified by: *Scott Als*

Analytical Services Chemist