

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	0382			
	AMMONIA	STANDARD 1,000 PPM AS	5 NH3 (AMMONIUM SULFATE)	
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES	-	YES
FILE		YES		
Sample ID				
Tested	02/03/25			
Tester	MMAT			
Recorded	02/03/25			
APPEARANCE AND COL	.OR	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	02/03/25			
Tester	MMAT			
Recorded	02/03/25			
INSTRUMENTS USED D	URING	INSTRUMENT		M108,VF1-Z31,M107
PREPARATION				
Sample ID				
Tested	02/03/25			
Tester	MMAT			
Recorded	02/03/25			
PPM NITROGEN AS NH3		1000	950 - 1050	1006
		PPM		
Sample ID				
Tested	02/03/25			
Tester	MMAT			
Recorded	02/03/25			
EXPIRATION DATE		MM/DD/YY		05/30/25
		MM/DD/YYYY		
Sample ID				
Tested	02/03/25			
Tester	MMAT			
Recorded	02/03/25			
Lot # 50201	77			
Made 02/03	/25			
induc -, ···				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA