



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

502019  
HA-26042-99

## Certificate of Analysis

**PRODUCT:** MolyVer® 1 Molybdenum Reagent

**PRODUCT NUMBER:** 2604299

**LOT NUMBER:** A4355

**MANUFACTURE DATE:** 01/07/2025

**DATE OF ANALYSIS:** 01/07/2025

TEST	SPECIFICATIONS	RESULTS
pH of a reagent blank	4 to 4.5	4.29
Absorbance of a 10.0 ppm Molybdenum Standard vs the reagent blank.	1.01 to 1.13 abs	1.113 abs
Reagent Blank: Measured at 365 nm in 1 inch cells.	0 to 0.022 abs	0.0140 abs

The expiration date is Dec 2029

Certified by: *Scott Als*

Analytical Services Chemist