



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

502004
HA-14548-28

Certificate of Analysis

PRODUCT: Citric Acid

PRODUCT NUMBER: 1454828

LOT NUMBER: A4198

MANUFACTURE DATE: 07/18/2024

DATE OF ANALYSIS: 07/18/2024

TEST	SPECIFICATIONS	RESULTS
Silica Function Test: Measured using a 1.0 ppm Silica and 10.0 ppm PO4 standard.	0.92 to 1.08 ppm	1.022 ppm

The expiration date is Jul 2029

Certified by: *Scott Als*

Analytical Services Chemist