



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 1500			
BUFFER REFERENCE pH 7.0/YELLOW			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 02/06/25			
Tester BPHI			
Recorded 02/06/25			
<b>PH OF PRODUCT/SOLUTION</b>	7.000 PH	6.990 - 7.010	6.993
Sample ID			
Tested 02/06/25			
Tester BPHI			
Recorded 02/06/25			
<b>+ PH Measurement Uncertainty</b>	0.014 PH	= 0.014	0.014
Sample ID			
Tested 02/06/25			
Tester BPHI			
Recorded 02/06/25			
<b>QCSP 20 PH CALIBRATION METHODS</b>	YES METHOD		YES
Sample ID			
Tested 02/06/25			
Tester BPHI			
Recorded 02/06/25			
<b>Actual Temp</b>	25.0 DEG C	24.9 - 25.1	25.0
Sample ID			
Tested 02/06/25			
Tester BPHI			
Recorded 02/06/25			
<b>pH Traceable to N.I.S.T. SRM'S</b>	YES YES		YES
Sample ID			
Tested 02/06/25			
Tester BPHI			
Recorded 02/06/25			
<b>PH NIST TRACEABLE TO SRM 185</b>	YES YES		YES
Sample ID			
Tested 02/06/25			
Tester BPHI			
Recorded 02/06/25			
<b>NIST TRACEABLE TO SRM186I AND 186II</b>	YES YES		YES
Sample ID			
Tested 02/06/25			
Tester BPHI			
Recorded 02/06/25			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>PH NIST TRACEABLE TO SRM 187</b>	YES YES	YES
Sample ID		
Tested	02/06/25	
Tester	BPHI	
Recorded	02/06/25	
<b>APPEARANCE AND COLOR</b>	<b>CLEAR YELLOW/ODORLESS APPEARANCE</b>	<b>PASS</b>
Sample ID		
Tested	02/06/25	
Tester	BPHI	
Recorded	02/06/25	
<b>INSTRUMENTS USED DURING PREPARATION</b>	<b>INSTRUMENT</b>	<b>M91,T-IX</b>
Sample ID		
Tested	02/06/25	
Tester	BPHI	
Recorded	02/06/25	
<b>EXPIRATION DATE</b>	<b>MM/DD/YY MM/DD/YYYY</b>	<b>02/28/27</b>
Sample ID		
Tested	02/06/25	
Tester	BPHI	
Recorded	02/06/25	
<b>SHELF LIFE OF PRODUCT</b>	<b>2 YEARS TIME</b>	<b>2 YEARS</b>
Sample ID		
Tested	02/06/25	
Tester	BPHI	
Recorded	02/06/25	
<b>Lot #</b>	5020559	
<b>Made</b>	02/06/25	
<b>Packaged Product:</b>	1500 BUFFER REFERENCE pH 7.0/YELLOW	
<b>Lot #</b>		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.