



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500601
HA-1041-99

Certificate of Analysis

PRODUCT: Molybdate Reagent for Silica

PRODUCT NUMBER: 104199

LOT NUMBER: A4270

MANUFACTURE DATE: 10/10/2024

DATE OF ANALYSIS: 10/10/2024

TEST	SPECIFICATIONS	RESULTS
Function Test: Absorbance of 50 ppm silica standard vs the reagent blank measured at 410 nm in 2 mm cells.	0.248 to 0.296 abs	0.2610 abs
Reagent Blank: Absorbance measured at 410 nm in 1 inch cells vs deionized water.	0 to 0.022 abs	0.0030 abs

The expiration date is Oct 2029

Certified by: *Scott Als*

Analytical Services Chemist