



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500626
HA-25895-99

Certificate of Analysis

This is a Component of 2589599 / LOT A4291

PRODUCT: Chlorophosphonazo Indicator Solution Pillows

PRODUCT NUMBER: 2589527

LOT NUMBER: A4291

MANUFACTURE DATE: 11/06/2024

DATE OF ANALYSIS: 11/11/2024

TEST	SPECIFICATIONS	RESULTS
The relative percent of the absorbance of a 0.25 ppm magnesium standard to the absorbance of the 0.50 ppm calcium standard.	82 to 118 %	86.9 %
Concentration of 0.50 mg/L standard	475 to 525 ppb	523.0 ppb
Difference in a CDTA treated reagent blank vs an untreated reagent blank.	-0.004 to 0.004 abs	0.0030 abs

The expiration date is Nov 2026

Certified by: *Scott Ala*

Analytical Services Chemist