

502020
HA-26043-99



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

Certificate of Analysis

PRODUCT: MolyVer® 2 Molybdenum Reagent

PRODUCT NUMBER: 2604399

LOT NUMBER: A4352

MANUFACTURE DATE: 01/07/2025

DATE OF ANALYSIS: 01/08/2025

TEST	SPECIFICATIONS	RESULTS
Absorbance of a 10.0 ppm Molybdenum standard vs the reagent blank.	1.01 to 1.13 abs	1.115 abs
Reagent Blank: Measured at 365 nm in a 1 inch cell.	0 to 0.022 abs	0.0040 abs

The expiration date is Dec 2029

Certified by: *Scott Als*

Analytical Services Chemist