



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500622

HA-22542-49

### Certificate of Analysis

**PRODUCT:** Citric Acid F Reagent

**PRODUCT NUMBER:** 2254249

**LOT NUMBER:** A4302

**MANUFACTURE DATE:** 11/06/2024

**DATE OF ANALYSIS:** 11/08/2024

TEST	SPECIFICATIONS	RESULTS
Silica Function Test: Concentration of a 1.0 ppm Silica standard vs a reagent blank.	0.92 to 1.08 ppm	1.002 ppm
Phosphate Blank: Absorbance of a 10.0 ppm phosphate standard vs a reagent blank.	0 to 0.015 abs	0.0020 abs
Silica Reagent Blank: A sample containing 2 mL of Citric Acid F Reagent in a Silica determination measured at 815 nm in 1 inch cells vs a sample containing 1 mL of reagent.	0 to 0.007 abs	0.0000 abs

The expiration date is Oct 2029

Certified by: *Scott Als*

Analytical Services Chemist