



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500602  
HA-1042-99

### Certificate of Analysis

**PRODUCT:** Acid Reagent

**PRODUCT NUMBER:** 104299

**LOT NUMBER:** A4321

**MANUFACTURE DATE:** 12/02/2024

**DATE OF ANALYSIS:** 12/02/2024

TEST	SPECIFICATIONS	RESULTS
Absorbance of a 150 ppm Silica standard vs the reagent blank in a 2 mm cell.	0.726 to 0.858 abs	0.7530 abs
Reagent Blank: Absorbance at 410 nm in a 1 inch cell vs DI water.	0 to 0.022 abs	0.0020 abs

The expiration date is Nov 2029

Certified by: *Scott Als*

Analytical Services Chemist