Made 02/03/25

Page: 1

Date: 02/03/25 at 3:10 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Customer PO#.					
Package	ed Product:	SPX475F			
•		NITROGEN	N STANDARD 1000 PPM	AS N AQUEOUS SOLUTION	
Test			Target/UOM	Range	Result
	D TO FORMUL	ATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	02/03/25			
	Tester	DTRA			
	Recorded	02/03/25			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS	
			LIQUID		
			APPEARANCE		
	Sample ID				
	•	02/03/25			
	Tester	DTRA			
	Recorded	02/03/25			
INSTRUM	ENTS USED DI	JRING	INSTRUMENT		M81,VF2-B12
PREPARA	TION				
	Sample ID				
	•	02/03/25			
		DTRA			
		02/03/25			
EXPIRATION DATE		MM/DD/YY		04/08/25	
			MM/DD/YYYY		
	Sample ID				
	•	02/03/25			
		DTRA			
	Recorded	02/03/25			
PPM NITROGEN NO DECIAML		1000	985 - 1015	991	
PLACES			PPM		
	Sample ID				
	-	02/03/25			
		DTRA			
		02/03/25			
Lo	ot # 502018	32			

Quality Manager