

500630
HA-26957-53



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

Certificate of Analysis

PRODUCT: APA Reagent 2 Hardness Buffer Solution

PRODUCT NUMBER: 2695753

LOT NUMBER: A4278

MANUFACTURE DATE: 10/14/2024

DATE OF ANALYSIS: 10/15/2024

TEST	SPECIFICATIONS	RESULTS
pH of the solution	10.3 to 10.7	10.67
Difference in absorbances measured in a one cm cells at 520 nm between a reagent blank containing EDTA and a reagent blank.	0 to 0.18 abs	0.136 abs

The expiration date is Oct 2025

Certified by: *Scott Ala*

Analytical Services Chemist