



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500616  
HA-21056-69

### Certificate of Analysis

**PRODUCT:** DPD Total Chlorine Reagent

**PRODUCT NUMBER:** 2105669

**LOT NUMBER:** A4331

**MANUFACTURE DATE:** 12/18/2024

**DATE OF ANALYSIS:** 12/18/2024

TEST	SPECIFICATIONS	RESULTS
pH of a reagent blank	6.2 to 6.5	6.38
Percent Recovery for a 2.5 ppm Standard. Chlorine concentration determined using DPD compared to the actual concentration.	93 to 107 %	96.3 %
Hardness Reagent Blank: Sample containing 1000 ppm as CaCO3 hardness vs DI water.	0 to 0.009 abs	0.0020 abs

The expiration date is Dec 2029

Certified by: *Scott Als*

Analytical Services Chemist