



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500618  
HA-21194-49

### Certificate of Analysis

**PRODUCT:** Nessler Reagent

**PRODUCT NUMBER:** 2119449

**LOT NUMBER:** A4331

**MANUFACTURE DATE:** 12/09/2024

**DATE OF ANALYSIS:** 12/06/2024

TEST	SPECIFICATIONS	RESULTS
Function Test: ABS of an Ammonia Nitrogen standard at concentration of 0.64 ppm NH4 as N measured at 460 nm in 1 inch cells vs DI water.	0.165 to 0.192 abs	0.1660 abs
Reagent Blank: Measured at 460 nm in an 1 inch cell vs DI water.	0 to 0.006 abs	0.0000 abs
Function Test: ABS of an Ammonia Nitrogen standard at concentration of 3.20 ppm NH4 as N measured at 460 nm in 1 inch cells vs DI water.	0.942 to 1.101 abs	0.9420 abs

The expiration date is Nov 2029

Certified by: *Scott Als*

Analytical Services Chemist