



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: MIL059F			
60% SODIUM PHOSPHATE 0.02 MOLAR 40%MEOH			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 01/27/25			
Tester DNEW			
Recorded 01/27/25			
PH OF PRODUCT/SOLUTION	3.000	2.995 - 3.005	2.997
	PH		
Sample ID			
Tested 01/27/25			
Tester DNEW			
Recorded 01/27/25			
FILTERED THROUGH A 0.2 MICRON FILTER	YES		YES
Sample ID	YES		
Tested 01/27/25			
Tester DNEW			
Recorded 01/27/25			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 01/27/25			
Tester DNEW			
Recorded 01/27/25			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		T-49,M108,3LVF3-J21,M106
Sample ID			
Tested 01/27/25			
Tester DNEW			
Recorded 01/27/25			
Lot #	5012445		
Made	01/27/25		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.