



Be Right

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500625  
HA-25895-99

### Certificate of Analysis

*This is a Component of 2589599 / LOT A4276*

**PRODUCT:** Chlorophosphonazo Indicator Solution Pillows

**PRODUCT NUMBER:** 2589527

**LOT NUMBER:** A4276

**MANUFACTURE DATE:** 10/30/2024

**DATE OF ANALYSIS:** 10/31/2024

TEST	SPECIFICATIONS	RESULTS
Concentration of 0.50 mg/L standard	475 to 525 ppb	500.0 ppb
Difference in a CDTA treated reagent blank vs an untreated reagent blank.	-0.004 to 0.004 abs	-0.0030 abs
The relative percent of the absorbance of a 0.25 ppm magnesium standard to the absorbance of the 0.50 ppm calcium standard.	82 to 118 %	86.4 %

The expiration date is Oct 2026

Certified by: *Scott Als*

Analytical Services Chemist