



Be Right™

Loveland, CO 80539

(970) 669-3050

500904  
HA-14548-28

An ISO 9001 Certified Company

### Certificate of Analysis

**PRODUCT:** Citric Acid

**PRODUCT NUMBER:** 1454828

**LOT NUMBER:** A4198

**MANUFACTURE DATE:** 07/18/2024

**DATE OF ANALYSIS:** 07/18/2024

TEST	SPECIFICATIONS	RESULTS
Silica Function Test: Measured using a 1.0 ppm Silica and 10.0 ppm PO4 standard.	0.92 to 1.08 ppm	1.022 ppm

The expiration date is Jul 2029

Certified by: *Scott Als*

Analytical Services Chemist