



Be Right™

Loveland, CO 80539

(970) 669-3050

503411

An ISO 9001 Certified Company

HA-23864-42

*Certificate of Analysis*

**PRODUCT:** Amino Acid F Reagent Solution

**PRODUCT NUMBER:** 2386442

**LOT NUMBER:** A4347

**MANUFACTURE DATE:** 12/30/2024

**DATE OF ANALYSIS:** 12/30/2024

TEST	SPECIFICATIONS	RESULTS
Absorbance of Silica standard vs the reagent blank with 2.0 ppm SiO <sub>2</sub> .	0.692 to 0.718 abs	0.6960 abs
Reagent Blank: Measured at 815 nm in 1 cm cell vs DI water.	0 to 0.005 abs	0.0011 abs
pH of the solution	7.1 to 7.7	7.57

The expiration date is Aug 2025

Certified by: *Scott Als*

Analytical Services Chemist