



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

502006  
HA-25895-49

### Certificate of Analysis

**PRODUCT:** Chlorophosponazo Indicator Solution

**PRODUCT NUMBER:** 2589549

**LOT NUMBER:** A4346

**MANUFACTURE DATE:** 12/18/2024

**DATE OF ANALYSIS:** 12/19/2024

TEST	SPECIFICATIONS	RESULTS
The relative percent of the absorbance of a 0.25 ppm magnesium standard to the absorbance of the 0.50 ppm calcium standard.	82 to 118 %	100.0 %
Concentration of 0.50 mg/L standard	475 to 525 ppb	475.0 ppb
Difference in a CDTA treated reagent blank vs an untreated reagent blank.	-0.004 to 0.004 abs	0.0010 abs

The expiration date is Dec 2026

Certified by: *Scott Als*

Analytical Services Chemist