



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500635
HA-67750-02

Certificate of Analysis

PRODUCT: 5500sc Standard, Silica 0.5 mg/L

PRODUCT NUMBER: 6775002

LOT NUMBER: A4304

MANUFACTURE DATE: 11/07/2024

DATE OF ANALYSIS: 11/07/2024

TEST	SPECIFICATIONS	RESULTS
Silica concentration.	0.495 to 0.505 mg/l	0.5030 mg/l

The expiration date is Nov 2026

The item 6775002 is traceable to NIST standards SRM 3150 Silicon 10 mg/mL LOT 130912

Certified by: *Scott Als*

Analytical Services Chemist