

500627
HA-23525-12



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

Certificate of Analysis

PRODUCT: Molybdenum ? Reagent for Low Range Molybdate

PRODUCT NUMBER: 2352512

LOT NUMBER: A4339

MANUFACTURE DATE: 12/05/2024

DATE OF ANALYSIS: 11/08/2024

TEST	SPECIFICATIONS	RESULTS
Function Test: Absorbance of 2.0 ppm molybdenum standard measured at 610 nm in a 1 inch cell.	0.941 to 1.013 abs	0.9730 abs

The expiration date is Apr 2028

Certified by: *Scott Als*

Analytical Services Chemist