

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: WES039F			
Brine Standard			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 02/10/25			
Tester MMAT			
Recorded 02/10/25			
CALCULATED PPM CALCIUM	1.00	0.99 - 1.01	1.00
	PPM		
Sample ID			
Tested 02/10/25			
Tester MMAT			
Recorded 02/10/25			
CALCULATED PPM MAGNESIUM (MG)	1.00	0.99 - 1.01	1.00
	PPM		
Sample ID			
Tested 02/10/25			
Tester MMAT			
Recorded 02/10/25			
CALCULATED PPM NICKEL	1.00	0.99 - 1.01	1.00
	PPM		
Sample ID			
Tested 02/10/25			
Tester MMAT			
Recorded 02/10/25			
CALCULATED PPM IRON (FE)	1.00	0.99 - 1.01	1.00
	PPM		
Sample ID			
Tested 02/10/25			
Tester MMAT			
Recorded 02/10/25			
CALCULATED MG/L SILICON (SI)	1.00	0.99 - 1.01	1.00
	MG/L		
Sample ID			
Tested 02/10/25			
Tester MMAT			
Recorded 02/10/25			
CALCULATED PPM SODIUM SULFATE	5000.0	4995.0 - 5005.0	4999.8
	PPM		
Sample ID			
Tested 02/10/25			
Tester MMAT			
Recorded 02/10/25			
APPEARANCE AND COLOR	PASS		PASS
	APPEARANCE		
Sample ID			
Tested 02/10/25			
Tester MMAT			
Recorded 02/10/25			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 02/10/25 at 2:16 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

INSTRUMENTS USED DURING

INSTRUMENT

P138,VF1-Z33,M108,GP

PREPARATION

Sample ID

Tested 02/10/25

Tester MMAT

Recorded 02/10/25

EXPIRATION DATE

MM/DD/YYYY

02/28/2026

MM/DD/YYYY

Sample ID

Tested 02/10/25

Tester MMAT

Recorded 02/10/25

Lot # 5020778

Made 02/10/25

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager