



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 1

Date: 02/26/25 at 6:26 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: SPE429F CALCIUM CARBONATE STANDARD 5000 PPM AS CaCO3			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION			
	YES		YES
	YES		
Sample ID			
Tested	02/24/25		
Tester	MMAT		
Recorded	02/24/25		
PPM AS CaCO3			
	5000	4990 - 5010	5000
	PPM		
Sample ID			
Tested	02/24/25		
Tester	MMAT		
Recorded	02/24/25		
APPEARANCE AND COLOR			
	CLEAR ORANGE		PASS
	APPEARANCE		
Sample ID			
Tested	02/24/25		
Tester	MMAT		
Recorded	02/24/25		
INSTRUMENTS USED DURING PREPARATION			
	LAB EQUIPMENT USED		M108,GP,VF1-Z32
	INSTRUMENT		
Sample ID			
Tested	02/24/25		
Tester	MMAT		
Recorded	02/24/25		
EXPIRATION DATE			
	MM/DD/YY		02/28/26.
	MM/DD/YYYY		
Sample ID			
Tested	02/24/25		
Tester	MMAT		
Recorded	02/24/25		
SHELF LIFE OF PRODUCT			
	SHELF LIFE		1 YEAR
	TIME		
Sample ID			
Tested	02/24/25		
Tester	MMAT		
Recorded	02/24/25		
Lot # 5022169			
Made 02/24/25			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 02/26/25 at 6:26 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager