02/10/25

Made

Page: 1

Date: 02/10/25 at 4:34 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged F	Product:	AM132F			
		CHLORIDE	STANDARD 10.0PPM W/W		
Test			Target/UOM	Range	Result
PREPARED TO	O FORMULA	ATION ON	YES		YES
FILE			YES		
S	ample ID				
Т	ested 0)2/10/25			
Т	ester J	JTRA			
R	Recorded C)2/10/25			
APPEARANCE AND COLOR)R	CLEAR WATER WHITE		PASS
			LIQUID		
			APPEARANCE		
s	ample ID				
		02/10/25			
Т		JTRA			
R	Recorded C)2/10/25			
INSTRUMENT		RING	INSTRUMENT		P46,P117,VF5-L12
PREPARATION	N				
s	ample ID				
		02/10/25			
Т		ITRA			
R)2/10/25			
PPM CHLORI			10.00	9.50 - 10.50	10.00
			PPM		
s	ample ID				
		02/10/25			
Т		ITRA			
R)2/10/25			
EXPIRATION DATE			MM/DD/YY		02/28/26
			MM/DD/YYYY		
s	ample ID				
		02/10/25			
Т		ITRA			
R	Recorded C)2/10/25			
Lot # 5020622					
LOT# 5020022					

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1.