

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8975			
SULFIDE STANDARD 1000 PPM W/W (1ML = 1MG S2-)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 02/25/25			
Tester JENC			
Recorded 02/25/25			
PPM SULFIDE TO 1 DECIMAL	1000.0	990.0 - 1010.0	999.6
PLACE	PPM		
Sample ID			
Tested 02/25/25			
Tester JENC			
Recorded 02/25/25			
QCSP56:STANDARDIZATION	YES		YES
SODIUM SULFIDE SO	QCSP56		
Sample ID			
Tested 02/25/25			
Tester JENC			
Recorded 02/25/25			
Actual Density	1.00608	1.00497 - 1.00719	1.00688
	G/ML		
Sample ID			
Tested 02/25/25			
Tester JENC			
Recorded 02/25/25			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 02/25/25			
Tester JENC			
Recorded 02/25/25			
INSTRUMENTS USED DURING	LAB EQUIPMENT		M112,VF12-99
PREPARATION	INSTRUMENT		
Sample ID			
Tested 02/25/25			
Tester JENC			
Recorded 02/25/25			
EXPIRATION DATE	MM/DD/YY		04/30/25
	MM/DD/YYYY		
Sample ID			
Tested 02/25/25			
Tester JENC			
Recorded 02/25/25			
SHELF LIFE OF PRODUCT	2 MONTHS		2 MONTHS
	TIME		
Sample ID			
Tested 02/25/25			
Tester JENC			
Recorded 02/25/25			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 02/25/25 at 10:04 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

PREPARED TO STANDARD

YES

YES

METHODS 4500-S2

Sample ID

Tested 02/25/25

Tester JENC

Recorded 02/25/25

Lot # 5022401

Made 02/25/25

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager