

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: ERL125F SILVER NITRATE 0.1 NORMAL IN ISOPROPANOL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		02/27/2025
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	19.0
SAMPLING DEGC	DEG C		
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		02/27/2025
	MM/DD/YYYY		
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	19.0
DEG C	DEG C		
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
Actual Normality	0.1000	0.0995 - 0.1005	0.1022
	EQ/L		
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
QCSP53: SILVER NITRATE STD ON	YES		YES
AUTO-TITRA	METHOD		
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 919 NaCl	YES		YES
	YES		
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
PREPARED TO ASTM D3227	YES		YES
	YES		
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
Actual Density	0.82561	0.82176 - 0.82947	0.82696
	G/ML		
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M55,P216,M52,TID53294
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
LOT NUMBER OF ORIGINAL MATERIAL	LOT #		4092307
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
EXPIRATION DATE	MM/DD/YY		03/11/25
	MM/DD/YYYY		
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
SHELF LIFE OF PRODUCT	1 WEEK		1 WEEK
	TIME		
Sample ID			
Tested 02/27/25			
Tester DTRA			
Recorded 02/27/25			
Lot #	5022418		
Made	02/27/25		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 02/27/25 at 2:20 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager