



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 7350			
POTASSIUM IODIDE-IODATE 0.1 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.1000 EQ/L	0.0995 - 0.1005	0.0999
Sample ID			
Tested	02/17/25		
Tester	LWIL		
Recorded	02/17/25		
Actual Density	1.02693 G/ML	1.02581 - 1.02805	1.02687
Sample ID			
Tested	02/17/25		
Tester	LWIL		
Recorded	02/17/25		
QCSP33:POTASSIUM IODATE-IODIDE TITRANT	YES QCSP33		YES
Sample ID			
Tested	02/17/25		
Tester	LWIL		
Recorded	02/17/25		
N.I.S.T. Traceable to SRM 136	Yes YES		YES
Sample ID			
Tested	02/17/25		
Tester	LWIL		
Recorded	02/17/25		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	02/17/25		
Tester	LWIL		
Recorded	02/17/25		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	02/17/25		
Tester	LWIL		
Recorded	02/17/25		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,T-165,M55,P362,BURET 00787
Sample ID			
Tested	02/17/25		
Tester	LWIL		
Recorded	02/17/25		
SHELF LIFE OF PRODUCT	1 YEAR TIME		1 YEAR
Sample ID			
Tested	02/17/25		
Tester	LWIL		
Recorded	02/17/25		



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

EXPIRATION DATE	MM/DD/YYYY	02/28/2026
	MM/DD/YYYY	
Sample ID		
Tested	02/17/25	
Tester	LWIL	
Recorded	02/17/25	
Lot #	5020702	
Made	02/17/25	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager