Page: 1

Date: 03/05/25 at 6:03 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	t: FIS096	F		
J		Standard 10.0 ppm in 2%	Nitric Acid Solution	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID)			
Tested	03/04/25			
Tester	CCOL			
Recorded	03/04/25			
PPM SODIUM		10.0	9.9 - 10.1	10.0
İ		PPM		
Sample ID)			
Tested	03/04/25			
Tester	CCOL			
Recorded	03/04/25			
APPEARANCE AND CO	OLOR	CLEAR WATER WHITE		PASS
		APPEARANCE		
Sample ID)			
Tested	03/04/25			
Tester	CCOL			
Recorded				
INSTRUMENTS USED		INSTRUMENT		P94,GC,VF1-Z29
PREPARATION				, ,
Sample ID)			
Tested	03/04/25			
Tester	CCOL			
Recorded				
EXPIRATION DATE	00/01/20	MM/DD/YYYY		03/30/2026
		MM/DD/YYYY		30,00,202
Sample ID)	,22,		
Tested	03/04/25			
Tester	CCOL			
Recorded	03/04/25			
Lot # 5030				
Made 03/0	1/40			

Quality Manager