

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 6740 PHENYLARSINE OXIDE TITRANT 0.00564N			
Test	Target/UOM	Range	Result
APPEARANCE AND COLOR CLEAR COLORLESS LIQUID APPEARANCE			
Sample ID			
Tested	03/05/25		
Tester	BPHI		
Recorded	03/05/25		
IDENTIFICATION PASSES TEST PASS			
Sample ID			
Tested	03/05/25		
Tester	BPHI		
Recorded	03/05/25		
PH OF PRODUCT/SOLUTION 6.000 PH			
		0.000 - 7.000	6.515
Sample ID			
Tested	03/05/25		
Tester	BPHI		
Recorded	03/05/25		
PREPARED TO FORMULATION ON FILE YES YES			
Sample ID			
Tested	03/05/25		
Tester	BPHI		
Recorded	03/05/25		
INSTRUMENTS USED DURING PREPARATION INSTRUMENT M112,M79,T-172,P62,M55,BUR ET06969			
Sample ID			
Tested	03/05/25		
Tester	BPHI		
Recorded	03/05/25		
NORMALITY TO 5 DECIMAL PLACES 0.00564 EQUIV/LITER			
		0.00559 - 0.00569	0.00560
Sample ID			
Tested	03/05/25		
Tester	BPHI		
Recorded	03/05/25		
SHELF LIFE OF PRODUCT 18 MONTHS TIME			
			18 MONTHS
Sample ID			
Tested	03/05/25		
Tester	BPHI		
Recorded	03/05/25		
Actual Density 0.99971 G/ML			
		0.99862 - 1.00081	1.00018
Sample ID			
Tested	03/05/25		
Tester	BPHI		
Recorded	03/05/25		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 03/06/25 at 7:42 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE

MM/DD/YY
MM/DD/YYYY

09/30/26

Sample ID

Tested 03/05/25

Tester BPHI

Recorded 03/05/25

Lot # 5022725

Made 03/05/25

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager