


Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8810			
SODIUM THIOSULFATE 0.1 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 03/05/25			
Tester DTRA			
Recorded 03/05/25			
Actual Normality	0.1000	0.0995 - 0.1005	0.0999
	EQ/L		
Sample ID			
Tested 03/05/25			
Tester DTRA			
Recorded 03/05/25			
QCSP10:STANDARDIZATION OF	YES		YES
SODIUM THIOSUL	QCSP10		
Sample ID			
Tested 03/05/25			
Tester DTRA			
Recorded 03/05/25			
TRACEABLE TO NIST SRM 136	YES		YES
	YES		
Sample ID			
Tested 03/05/25			
Tester DTRA			
Recorded 03/05/25			
Actual Density	1.01123	1.01079 - 1.01166	1.01142
	G/ML		
Sample ID			
Tested 03/05/25			
Tester DTRA			
Recorded 03/05/25			
PH OF PRODUCT/SOLUTION	7.000	6.000 - 12.500	10.220
	PH		
Sample ID			
Tested 03/05/25			
Tester DTRA			
Recorded 03/05/25			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 03/05/25			
Tester DTRA			
Recorded 03/05/25			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		03/05/2025
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 03/05/25			
Tester DTRA			
Recorded 03/05/25			

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	19.8
Sample ID			
Tested	03/05/25		
Tester	DTRA		
Recorded	03/05/25		
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		03/05/2025
Sample ID			
Tested	03/05/25		
Tester	DTRA		
Recorded	03/05/25		
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	20.0
Sample ID			
Tested	03/05/25		
Tester	DTRA		
Recorded	03/05/25		
Additions, Deviations or Exclusions	NONE CHARACTER		NONE
Sample ID			
Tested	03/05/25		
Tester	DTRA		
Recorded	03/05/25		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	03/05/25		
Tester	DTRA		
Recorded	03/05/25		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M103,M112,T72,M55
Sample ID			
Tested	03/05/25		
Tester	DTRA		
Recorded	03/05/25		
DATE OF PREPARATION	MM/DD/YYYY		03/05/2025
Sample ID			
Tested	03/05/25		
Tester	DTRA		
Recorded	03/05/25		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		09/30/25
Sample ID			
Tested	03/05/25		
Tester	DTRA		
Recorded	03/05/25		



SHELF LIFE		6 MONTHS	6 MONTHS
		TIME	
Sample ID			
Tested	03/05/25		
Tester	DTRA		
Recorded	03/05/25		
Lot #	5022512		
Made	03/05/25		

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Richard E. Roth

Quality Manager