



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 1

Date: 03/04/25 at 6:32 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: FIS097F 10 ppm Pb in 2% Nitric Acid Solution			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 03/03/25			
Tester CCOL			
Recorded 03/03/25			
PPM LEAD (PB)	10.0 PPM	9.9 - 10.1	10.0
Sample ID			
Tested 03/03/25			
Tester CCOL			
Recorded 03/03/25			
APPEARANCE AND COLOR	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 03/03/25			
Tester CCOL			
Recorded 03/03/25			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P83,GC,VF500-X41
Sample ID			
Tested 03/03/25			
Tester CCOL			
Recorded 03/03/25			
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		03/30/2026
Sample ID			
Tested 03/03/25			
Tester CCOL			
Recorded 03/03/25			
Lot #	5030364		
Made	03/03/25		

Quality Manager