

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 209861F			
Sulfuric Acid, Standardized 0.02N (N/50) (BCG/MR)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 03/01/25			
Tester JENC			
Recorded 03/01/25			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 03/01/25			
Tester JENC			
Recorded 03/01/25			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		03/01/2025
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 03/01/25			
Tester JENC			
Recorded 03/01/25			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	20.3
SAMPLING DEGC	DEG C		
Sample ID			
Tested 03/01/25			
Tester JENC			
Recorded 03/01/25			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		03/01/2025
	MM/DD/YYYY		
Sample ID			
Tested 03/01/25			
Tester JENC			
Recorded 03/01/25			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	20.3
DEG C	DEG C		
Sample ID			
Tested 03/01/25			
Tester JENC			
Recorded 03/01/25			
NORMALITY TO 4 DEC PLACES	0.0200	0.0195 - 0.0205	0.0200
	EQUIV/LITER		
Sample ID			
Tested 03/01/25			
Tester JENC			
Recorded 03/01/25			
QCSP61A:STANDARDIZATION	YES		YES
ACIDS USING SODI	QCSP61A		
Sample ID			
Tested 03/01/25			
Tester JENC			
Recorded 03/01/25			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 03/03/25 at 6:42 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

APPEARANCE AND COLOR

CLEAR LIQUID
APPEARANCE

PASS

Sample ID
Tested 03/01/25
Tester JENC
Recorded 03/01/25

**INSTRUMENTS USED DURING
PREPARATION**

INSTRUMENT

GC,T-579.2895,M55,M91,BUR
ET06419

Sample ID
Tested 03/01/25
Tester JENC
Recorded 03/01/25

EXPIRATION DATE

MM/DD/YYYY
MM/DD/YYYY

03/30/2027

Sample ID
Tested 03/01/25
Tester JENC
Recorded 03/01/25

Lot # 5022006

Made 03/01/25

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager