



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 1

Date: 03/04/25 at 6:35 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: ND229F TOC 3000 MG/L & TIC 200 MG/L			
Test	Target/UOM	Range	Result
APPEARANCE AND COLOR	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested	03/03/25		
Tester	DTRA		
Recorded	03/03/25		
CERTIFIED TOC VALUE	3000 PPM TOC	2900 - 3100	2979
Sample ID			
Tested	03/03/25		
Tester	DTRA		
Recorded	03/03/25		
CERTIFIED IC VALUE	200 ACT IC	185 - 215	186
Sample ID			
Tested	03/03/25		
Tester	DTRA		
Recorded	03/03/25		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M81,P101,VF1-Z35,M66,M109
Sample ID			
Tested	03/03/25		
Tester	DTRA		
Recorded	03/03/25		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		05/08/25
Sample ID			
Tested	03/03/25		
Tester	DTRA		
Recorded	03/03/25		
Lot #	5022848		
Made	03/03/25		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.