



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9563 TOTAL ORGANIC CARBON STANDARD 5.0 mg/L (ppm)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON YES YES			
FILE YES			
Sample ID			
Tested 03/11/25			
Tester DTRA			
Recorded 03/11/25			
PRESERVED WITH HYDROCHLORIC YES YES			
ACID YES			
Sample ID			
Tested 03/11/25			
Tester DTRA			
Recorded 03/11/25			
PPM TOC 5.00 4.70 - 5.30 4.86			
PPM			
Sample ID			
Tested 03/11/25			
Tester DTRA			
Recorded 03/11/25			
APPEARANCE AND COLOR CLEAR WATER WHITE LIQUID APPEARANCE PASS			
Sample ID			
Tested 03/11/25			
Tester DTRA			
Recorded 03/11/25			
INSTRUMENTS USED DURING INSTRUMENT P94,VF2-B11,M109			
PREPARATION			
Sample ID			
Tested 03/11/25			
Tester DTRA			
Recorded 03/11/25			
EXPIRATION DATE MM/DD/YY 05/16/2025			
MM/DD/YYYY			
Sample ID			
Tested 03/11/25			
Tester DTRA			
Recorded 03/11/25			
SHELF LIFE OF PRODUCT SHELF LIFE 2 MONTHS			
TIME			
Sample ID			
Tested 03/11/25			
Tester DTRA			
Recorded 03/11/25			
Lot # 5030509			
Made 03/11/25			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 03/12/25 at 11:36 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager