



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8754 SODIUM STANDARD 1.0 PPM (1000PPB) NIST TRACEABLE SOLUTION			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 03/06/25			
Tester DTRA			
Recorded 03/06/25			
Calculated PPM SODIUM	1.00 PPM	0.98 - 1.02	1.00
Sample ID			
Tested 03/06/25			
Tester DTRA			
Recorded 03/06/25			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 03/06/25			
Tester DTRA			
Recorded 03/06/25			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P94,P160,VF12-101
Sample ID			
Tested 03/06/25			
Tester DTRA			
Recorded 03/06/25			
SHELF LIFE OF PRODUCT	6 MONTHS TIME		6 MONTHS
Sample ID			
Tested 03/06/25			
Tester CCOL			
Recorded 03/06/25			
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		09/30/2025
Sample ID			
Tested 03/06/25			
Tester CCOL			
Recorded 03/06/25			
Lot #	5030618		
Made	03/06/25		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 03/11/25 at 7:02 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Quality Manager