

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8580 SODIUM CARBONATE 0.05 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.0500 EQ/L	0.0495 - 0.0505	0.0502
Sample ID			
Tested 03/11/25			
Tester BPHI			
Recorded 03/11/25			
STANDARDIZED USING QCSP36	YES PASS		YES
Sample ID			
Tested 03/11/25			
Tester BPHI			
Recorded 03/11/25			
N.I.S.T. Traceable to SRM 723	Yes YES		YES
Sample ID			
Tested 03/11/25			
Tester BPHI			
Recorded 03/11/25			
Actual Density	1.00098 G/ML	1.00087 - 1.00108	1.00103
Sample ID			
Tested 03/11/25			
Tester BPHI			
Recorded 03/11/25			
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 03/11/25			
Tester BPHI			
Recorded 03/11/25			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 03/11/25			
Tester BPHI			
Recorded 03/11/25			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M103,GC,T-14,P393,BURET08 449
Sample ID			
Tested 03/11/25			
Tester BPHI			
Recorded 03/11/25			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		03/30/26
Sample ID			
Tested 03/11/25			
Tester BPHI			
Recorded 03/11/25			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 03/12/25 at 11:30 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

1 YEAR
TIME

1 YEAR

Sample ID

Tested 03/11/25

Tester BPHI

Recorded 03/11/25

Lot # 5030728

Made 03/11/25

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager