



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: VUL198F			
40% Sodium Acetate 0.1Molar in 60% v/v Methanol (Make fresh)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON			
	YES		YES
FILE			
	YES		
Sample ID			
Tested	03/14/25		
Tester	CCOL		
Recorded	03/14/25		
APPEARANCE AND COLOR			
	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested	03/14/25		
Tester	CCOL		
Recorded	03/14/25		
INSTRUMENTS USED DURING PREPARATION			
	INSTRUMENT		VF4-I5,VF4-N9,VF6-28,T23
Sample ID			
Tested	03/14/25		
Tester	CCOL		
Recorded	03/14/25		
EXPIRATION DATE			
	MM/DD/YYYY		07/30/2025
	MM/DD/YYYY		
Sample ID			
Tested	03/14/25		
Tester	CCOL		
Recorded	03/14/25		
Lot #	5031260		
Made	03/14/25		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager