

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9148 SULFURIC ACID 1.8 NORMAL W/0.008% MANGANESE SULFATE			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE Sample ID Tested 03/11/25 Tester BPHI Recorded 03/11/25	PASS YES		PASS
SAMPLE COLLECTION & HANDLING PER LAB040 Sample ID Tested 03/11/25 Tester BPHI Recorded 03/11/25	YES YES		YES
DATE OF SAMPLING FOR TEST PURPOSES Sample ID Tested 03/11/25 Tester BPHI Recorded 03/11/25	MM/DD/YYYY MM/DD/YYYY		03/11/2025
AMBIENT TEMPERATURE DURING SAMPLING DEGC Sample ID Tested 03/11/25 Tester BPHI Recorded 03/11/25	25.0 DEG C	10.0 - 35.0	20.0
DATE OF TESTING OF SAMPLE Sample ID Tested 03/11/25 Tester BPHI Recorded 03/11/25	MM/DD/YYYY MM/DD/YYYY		03/11/2025
AMBIENT TEST TEMPERATURE DEG C Sample ID Tested 03/11/25 Tester BPHI Recorded 03/11/25	20.0 DEG C	15.0 - 30.0	20.0
NORMALITY TO 4 DEC PLACES Sample ID Tested 03/11/25 Tester BPHI Recorded 03/11/25	1.8000 EQUIV/LITER	1.7991 - 1.8009	1.7998
SUITABLE FOR TOC USE Sample ID Tested 03/11/25 Tester BPHI Recorded 03/11/25	PASS PASS		PASS



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 03/12/25 at 11:32 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

APPEARANCE AND COLOR

CLEAR WATER
WHITE/ODORLESS
APPEARANCE

PASS

Sample ID
Tested 03/11/25
Tester BPHI
Recorded 03/11/25

INSTRUMENTS USED DURING PREPARATION

INSTRUMENT

T-A7,T-212,M112,M55,P106,B
URET0844

Sample ID
Tested 03/11/25
Tester BPHI
Recorded 03/11/25

Actual Density

1.05461
G/ML

1.05421 - 1.05501

1.05479

Sample ID
Tested 03/11/25
Tester BPHI
Recorded 03/11/25

EXPIRATION DATE

MM/DD/YY
MM/DD/YYYY

03/30/26

Sample ID
Tested 03/11/25
Tester BPHI
Recorded 03/11/25

Lot # 5030537

Made 03/11/25

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager