

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: SPE432			
WATER, DIUF (Ultra-filtered) (20L Cube) (FIS#W2-20)			
Test	Target/UOM	Range	Result
APPEARANCE AND COLOR	CLEAR COLORLESS LIQUID APPEARANCE		CLEAR COLORLESS LIQUID
Sample ID			
Tested 02/03/25			
Tester TBUR			
Recorded 02/03/25			
COLOR	< 5 PT-CO		<5 APHA
Sample ID			
Tested 02/03/25			
Tester TBUR			
Recorded 02/03/25			
IDENTIFICATION	PASS/FAIL PASS		PASS
Sample ID			
Tested 02/03/25			
Tester TBUR			
Recorded 02/03/25			
PH OF PRODUCT/SOLUTION	5.800 PH	5.400 - 7.000	5.800
Sample ID			
Tested 02/03/25			
Tester TBUR			
Recorded 02/03/25			
SPECIFIC CONDUCTANCE	< 2.0		0.9
Sample ID			
Tested 02/03/25			
Tester TBUR			
Recorded 02/03/25			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		11/07/2029
Sample ID			
Tested 02/03/25			
Tester TBUR			
Recorded 02/03/25			
SHELF LIFE OF PRODUCT	SHELF LIFE TIME		5 YEARS
Sample ID			
Tested 02/03/25			
Tester TBUR			
Recorded 02/03/25			
MANUFACTURE DATE	MM/DD/YY MM/DD/YYYY		11/07/2024
Sample ID			
Tested 02/03/25			
Tester TBUR			
Recorded 02/03/25			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 02/03/25 at 10:28 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Lot #	502904
Made	01/29/25

Quality Manager