



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: OXY203F			
LACHAT CALIBRATION STANDARD LEVEL 3			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 03/10/25			
Tester MMAT			
Recorded 03/10/25			
APPEARANCE AND COLOR	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 03/10/25			
Tester MMAT			
Recorded 03/10/25			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P58,VF2-B4,M120,M108,M55
Sample ID			
Tested 03/10/25			
Tester MMAT			
Recorded 03/10/25			
Actual Density	1.21729 G/ML	1.21501 - 1.21956	1.21607
Sample ID			
Tested 03/10/25			
Tester MMAT			
Recorded 03/10/25			
PPM KIO3-NAOCL	40.0 PPM	39.9 - 40.1	40.0
Sample ID			
Tested 03/10/25			
Tester MMAT			
Recorded 03/10/25			
CALCULATED GPL SODIUM HYDROXIDE	150.00 GPL	149.95 - 150.05	150.00
Sample ID			
Tested 03/10/25			
Tester MMAT			
Recorded 03/10/25			
CALCULATED GPL SODIUM CHLORIDE	130.00 GPL	129.98 - 130.02	130.00
Sample ID			
Tested 03/10/25			
Tester MMAT			
Recorded 03/10/25			
CALCULATED GPL SODIUM CHLORATE	1.0000 GPL	0.9995 - 1.0005	1.0002
Sample ID			
Tested 03/10/25			
Tester MMAT			
Recorded 03/10/25			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

1 YEAR
 TIME

1 YEAR

Sample ID

Tested 03/10/25

Tester MMAT

Recorded 03/10/25

EXPIRATION DATE

MM/DD/YYYY
 MM/DD/YYYY

03/30/2026

Sample ID

Tested 03/10/25

Tester MMAT

Recorded 03/10/25

Lot # 5031053

Made 03/10/25

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager