



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 1

Date: 03/14/25 at 2:45 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: M7200 MINERAL SPIRITS ODORLESS RULE 66			
Test	Target/UOM	Range	Result
GRAVITY, API @ 60 DEC F	MAX 55.5 DEG F		53.3
Sample ID			
Tested 03/14/25			
Tester TBUR			
Recorded 03/14/25			
FLASH PT, TCC	MIN 38(100) C(F) DEG F		43(110)
Sample ID			
Tested 03/14/25			
Tester TBUR			
Recorded 03/14/25			
SAYBOLT COLOR	MIN 28 COLOR		30
Sample ID			
Tested 03/14/25			
Tester TBUR			
Recorded 03/14/25			
AROMATICS, VOL%	MAX 1.0% AROMATIC%		PASS
Sample ID			
Tested 03/14/25			
Tester TBUR			
Recorded 03/14/25			
WATER CONTENT	MAX 500 PPM %		50 PPM
Sample ID			
Tested 03/14/25			
Tester TBUR			
Recorded 03/14/25			
MANUFACTURE DATE	MM/DD/YYYY MM/DD/YYYY		09/17/2024
Sample ID			
Tested 03/14/25			
Tester TBUR			
Recorded 03/14/25			
Lot #	505902		
Made	02/28/25		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 03/14/25 at 2:45 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager