



Aqua Solutions, Inc.

P.O. Box 70 Deer Park, TX 77536 Ph (281-479-2569)



SPECIFICATION SHEET

EDTA TITRANT 0.01 MOLAR (0.02 NORMAL) NIST TRACEABLE

Product Code	3550
Product Name	EDTA Titrant 0.01 Molar (0.02 Normal) NIST Traceable
Actual Molarity	0.0100 ± 0.0005 Molar @ 20°C
N.I.S.T Traceable	Yes
Prepared to Formulation on File	Yes
Appearance and Color	Clear Water White Liquid
Country of Origin	USA
Harmonization Code	2922.49.8000

The weights and /or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All Volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

PREPARATION DATE (MM/DD/YY)

Original 0.0	04/13/2018: STN
Revision	

PLEASE REFER TO THE MATERIAL SAFETY DATA SHEET FOR HEALTH, SAFETY, AND REGULATORY INFORMATION ON THIS PRODUCT.